

Indira Gandhi Medical College & Research Institute

(Government of Puducherry Institution)

Vazhudavur Road, Kathirkamam, Puducherry - 605 009 Ph :0413-2277545, Fax:0413-2277289

Email: puducherrymedicalcollege@gmail.com, Web:www.igmcri.com

No. 38/IGMC&RI/Estt-Pur/A5/ Bioche-Equip /19-20/1098 Date: -8 APR 2019

Sub: IGMC&RI – Quotation Call letter for Semi Automated Coagulation Analyser - Reg.

* * *


The following items are proposed to be purchased **Semi Automated Coagulation Analyser (for PT assay) for Biochemistry Lab** in this Institution. It is therefore requested to quote the lowest possible rate for the supply of the following with terms and conditions here under. The **Hard Copy** of the quotation shall be submitted as to **the Director, IGMC&RI, Puducherry on or before 26.04.19**

Sl.No.	Name of Item	Qty
1	Semi Automated Coagulation Analyser (for PT assay)	1 No..
Technical specifications enclosed separately		

Terms and conditions:

1. The supplier should be responsible until the items arrived to this Institution in good condition
2. The rates once quoted and approved will be final and no revision in rates will be allowed
3. Supply should be made within 15 days from the date of receipt of our supply order
4. Delivery should be made on FOR, Puducherry, door delivery basis.
5. The quotation should be based strictly on the specification.
6. **Prices should be inclusive of all taxes.**
7. The material should be supplied strictly in accordance with the specifications enclosed/ mentioned. The items which are not conforming to the specifications should be taken back at their cost.

Yours faithfully,


DIRECTOR
Indira Gandhi Medical College
& Research Institute
PUDUCHERRY,

To:

1. The Asst. Programmer, IGMC&RI, Puducherry- with instruction to Upload in our official Web-Site and for compliance report.

Copy to:


1. The HOD (Biochemistry), IGMC&RI., Puducherry.

INDIRA GANDHI MEDICAL COLLEGE & RESEARCH INSTITUTE
(GOVERNMENT OF PUDUCHERRY INSTITUTION)
DEPARTMENT OF BIOCHEMISTRY

Specification for Semi-automated, Single Channel Coagulation Analyser for
Prothrombin Time (PT)

- Open system for clot- reagent
- Turbidensitometric principle
- Test sample – plasma
- Single Channel for Prothrombin Time (PT)
- Results in Second & INR
- ISI values Editable
- Digital display of Results
- Inbuilt Printer to print the results
- Memory: storage of minimum 500 results
- Cuvette incubation position : Minimum 5
- Reagent incubation position : Minimum 2
- Sample volume 20-50 micro lit
- Reagent volume 40-100 micro lit
- Curvette Volume for assay - maximum 150 micro lit
- Incubation temperature: $37.0 \pm 1.0^{\circ} \text{c}$
- Power supply: AC 100 -240V, 50/60Hz
- Following accessories must be supplied
 - a) Cuvette at least 1000pcs
 - b) steel balls at least 1000pcs
 - c) Ejector for steel balls: one unit
 - d) Micropipette : 20 to 200 micro lit – at least one
 - UPS with minimum 30minutes backup
 - Free Installation & Training.

Warranty period 'Minimum Three years'.



25/03/2019